

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;">Accredited to ISO/IEC 17025:2005</p>	<h3 style="margin: 0;">HFS Environmental Ltd</h3> <p style="margin: 0;">Issue No: 003 Issue date: 25 January 2008</p>	
	<p>Lockheed House Green Lane Business Park New Eltham SE9 3TL</p>	<p>Contact: Mr C Healey Tel: +44 (0)20 8859 6633 Fax: E-Mail: lab@hfsenvironmental.co.uk Website: www.hfsenvironmental.co.uk</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Lockheed House Green Lane Business Park New Eltham SE9 3TL</p> <p>Contact: Mr C Healey Tel: +44 (0)20 8859 6633 Fax: Email: lab@hfsenvironmental.co.uk Website: www.hfsenvironmental.co.uk</p>	Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



4026
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

HFS Environmental Ltd
Issue No: 003 Issue date: 25 January 2008

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	A, B, C
	Fibre counting	HSG 248:February 2005 by Documented In-House Procedure TTP 5.3, Membrane Filter Method using Phase Contrast Microscopy (PCM)	
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure TTP 5.2 using stereo-microscopy, polarised light microscopy and dispersion staining	A
END			